## Nixie Tube Power Supply Test Report

Wilhelm Zeuschner

Report date: October 2023

## 1 Methodology

Each sold device has been tested individually before shipping using an automated process. The test setup is shown in figure 1.

The test results are written to a file and included inside the package for each sold device.

In case of failure the module is not to be sold and may be repaired prior to re-testing and sale.

Each module has to pass the tests specified below.



Figure 1: Automated test setup using test jig.

During the test the voltages of the different supply rails are measured. Since a fixed resistor is used as a load per rail, no current measurement is necessary.

## Test procedure:

- 1. Verify that the 3.3 V and 5 V rail are within spec under no load.
- 2. Verify that the high voltage is within spec under no load.
- 3. Verify that the 5 V rail is within spec under 1.5 A load.
- 4. Verify that the 3.3 V and 5 V rail are within spec under 0.44 A / 1.5 A load (both rails are loaded simultaneously).
- 5. Verify that the high voltage rail is within spec under 12 mA load.
- 6. Verify that the 3.3 V, 5 V and high voltage rail are within spec under 0.44 A / 1.5 A / 10 mA load (all rails are loaded simultaneously).

## Allowed deviation of measured values during test from calculated ideal values:

- 3.3 V: +/- 6%
- 5 V: +/- 6%
- $\bullet$  High voltage: +/- 10%

Please note that the output voltage of the high voltage boost converter can be calibrated since it is adjustable. The test script is not able to do so and thus a relatively large margin of  $\pm 10\%$  was set.

Thus, the margin of +/- 10% does not accurately reflect real-world accuracy which can be better with proper calibration.

-- Nixie HV PSU Module Tester --Firmmare Version: 1.10 Firmmare Release Date: 92. Nov. 2021 Author: Wilhelm Zeuschner https://wizeus.de/home/projects/adj-hv-psu/ (c) Wilhelm Zeuschner 2021, all rights reserved Compiled on: 0 ct 6 2023, 22:44:45 Starting tests... TEST STEP: MEASUREMENTS: 5.05 5V: 3.3V: PASS: 5V output voltage within spec. 3.3V output voltage within spec. PASS: STEP 1 DONE WITH NO ERRORS Verifying high voltage is actually present. Verifying high voltage output accuracy. SET VALUE: 82. 82.58 MEAS VALUE: HV output accuracy OK for this step! 87.21 85.97 SET VALUE: MEAS VALUE: PASS: HV output accuracy OK for this step! SET VALUE 92.41 91.71 HV output accuracy OK for this step! MEAS VALUE: PASS: SET VALUE: MEAS VALUE: HV output accuracy OK for this step! SET VALUE: MEAS VALUE: PASS: 104.95 104.58 HV output accuracy OK for this step! SET VALUE 112.61 112.63 HV output accuracy OK for this step! MEAS VALUE: PASS: SET VALUE: MEAS VALUE: PASS: 121.50 122.05 HV output accuracy OK for this step! MEAS VALUE: PASS: 132.86 HV output accuracy OK for this step! SET VALUE: MEAS VALUE: PASS: 144.36 145.73 HV output accuracy OK for this step! SET VALUE: 159.39 162.51 MEAS VALUE: PASS: HV output accuracy OK for this step! SET VALUE: MEAS VALUE: PASS: 177.97 171.01 HV output accuracy OK for this step! STEP 2 DONE WITH NO ERRORS TEST STEP: MEASUREMENTS before Load on: 5V: Starting periodic measurements. Reporting current value every 5 seconds. 5.07V 5.07V SV MEASUREMENTS after: 5s: 5V MEASUREMENTS after: 10s: 5V MEASUREMENTS after: 15s: 5V MEASUREMENTS after: 20s: 5.06V 5.08V 5V MEASUREMENTS after: 25s: 5V MEASUREMENTS after: 30s: 5V MEASUREMENTS after: 35s: 5.08V 5.07V 5.07V 5V MEASUREMENTS after: 40s: 5V MEASUREMENTS after: 45s: 5V MEASUREMENTS after: 50s: 5.07V 5.06V 5.06V 5V MEASUREMENTS after: 55s: 5V MEASUREMENTS after: 60s: 5.07V 5.07V 5V output voltage within spec. PASS: STEP 3 DONE WITH NO ERRORS 5.28 S.30: Starting periodic measurements. Reporting measured value every 5 seconds 3.3V MEASUREMENTS after: 5s: 3.38V 5V MEASUREMENT: 5.64V 5V MEASUREMENT: PASS: 5V output voltage within spec PASS: 3.3V MEASUREMENTS after: 10s: 5V MEASUREMENT: 5V output voltage within spec. 3.3V output voltage within spec PASS: 3.3V MEASUREMENTS after: 15s: 5V MEASUREMENT: PASS: PASS: PASS: 5V MEASUREMENTS after: 20s: 5V MEASUREMENT: PASS: PASS: 5V MEASUREMENT: PASS: 5V MEASUREMENTS after: 25s: 5V MEASUREMENT: PASS: 5V output voltage within spec. 3.3V output voltage within spec 3.38V 5.04V output voltage within spec. 3.38V 5.04V output voltage within spec. 3.38V 5.04V output voltage within spec. 3.3V output voltage within spec. 3.3V output voltage within spec. 3.3V output voltage within spec. PASS: PASS: 3.3V MEASUREMENTS after: 30s: 5V MEASUREMENT: 3.3V output voltage within spec. 3.38V 5.03V

5V output voltage within spec. 3.3V output voltage within spec.

5.38V 5.04V 5V output voltage within spec. 3.3V output voltage within spec.

PASS: PASS: 3.3V MEASUREMENTS after: 35s:

PASS: 3.3V MEASUREMENTS after: 40s: 5V MEASUREMENT:

5V MEASUREMENT: PASS: PASS: PASS: 3.3V MEASUREMENTS after: 45s: SV MEASUREMENT: PASS: 5V output voltage within spec. 3.3V output voltage within spec 3.38V output voltage within spec. 3.38V 5.64V
5.04V within spec. 3.3V output voltage within spec. 3.38V PASS: 3.3V MEASUREMENTS after: 50s 5V MEASUREMENT: 5.04V
5V output voltage within spec.
3.3V output voltage within spec.
3.38V
5.04V
5.04V
5.04V
5.04V
5.04V
6.04V
6. PASS: PASS: 3.3V MEASUREMENTS after: 55s 5V MEASUREMENT: PASS: PASS: STEP 4 DONE WITH NO ERRORS TEST STEP: Starting tests for HV.
Verifying high voltage is actually present. Verifying high voltage output accuracy under load. SET VALUE: 82.58 MEAS VALUE: 81.37 HV output accuracy OK for this sten! PASS: SET VALUE: 87.21 MEAS VALUE: 86.66 HV output accuracy OK for this step! SET VALUE: 92.41 91.94 MEAS VALUE: HV output accuracy OK for this step! SET VALUE 98 27 98.15 HV output accuracy OK for this step! MEAS VALUE: PASS: SET VALUE: MEAS VALUE: PASS: HV output accuracy OK for this step! SET VALUE: MEAS VALUE: PASS: 112.61 112.86 HV output accuracy OK for this step! SET VALUE: MEAS VALUE: PASS: 121.50 122.28 HV output accuracy OK for this step! SET VALUE: MEAS VALUE: PASS: 132.40 HV output accuracy OK for this step! MEAS VALUE: PASS: 145.96 HV output accuracy OK for this step! STEP 5 DONE WITH NO ERRORS 

5V: 5.28
3.39
Verifying high voltage is actually present
Starting periodic measurements. ReportLine measured value every 5 seconds.
3.3V MEASUREMENTS after: 5s: 3.3VS VS MEASUREMENTS after: 5s: 5.84V
PASS: 5.84V and measures value every becomes.
3.39V
5.40M tput voltage within spec.
3.3V output voltage within spec.
128.62
129.64
149 output accuracy OK for this step! PASS: PASS: SET VALUE: MEAS VALUE: 3.3V MEASUREMENTS after: 10s 5V MEASUREMENT: 5.03V 5V output voltage within spec. PASS: PASS: SET VALUE: 3.3V output voltage within spec. 128.62 MEAS VALUE: 129.41 HV output accuracy OK for this sten! PASS: 3.3V MEASUREMENTS after: 15s: 3.39V 5.04V 5V output voltage within spec. 3.3V output voltage within spec. 5V MEASUREMENT: PASS: PASS: SET VALUE: MEAS VALUE: 128.62 129.64 HV output accuracy OK for this step! 3.3V MEASUREMENTS after: 20s 3.39V 5V MEASUREMENT 5.04V PASS: PASS: SET VALUE: 5V output voltage within spec. 3.3V output voltage within spec. 128.62 MEAS VALUE: 129.87 HV output accuracy OK for this step! 3.3V MEASUREMENTS after: 25s 3.39V 5.04V PASS: PASS: SET VALUE: MEAS VALUE: PASS: 5.04V 5V output voltage within spec. 3.3V output voltage within spec. 128.62 129.64 HV output accuracy OK for this step! 3.39V 5.85V 5V output voltage within spec. 3.3V output voltage within spec. 128.62 3.3V MEASUREMENTS after: 30s: MEAS VALUE: PASS: 130.10 HV output accuracy OK for this step! 3.3V MEASUREMENTS after: 35s: 5V MEASUREMENT: 3.39V 5.04V 5V output voltage within spec. PASS: PASS: SET VALUE: 3.3V output voltage within spec. 128.62 MEAS VALUE: 129.87 HV output accuracy OK for this step! 3.3V MEASUREMENTS after: 40s: 5V MEASUREMENT: PASS: 5.04V 5V output voltage within spec.

PASS: SET VALUE: MEAS VALUE: PASS: 3.3V output voltage within spec. 128.62 129.41 HV output accuracy OK for this step! 3.3V MEASUREMENTS after: 45s: 5V MEASUREMENT: PASS: PASS: SET VALUE: MEAS VALUE: DASC: 3.39V 5.65V output voltage within spec. 3.3V output voltage within spec. 128.62 129.64 HV output accuracy OK for this step! 3.3V MEASUREMENTS after: 50s: 5V MEASUREMENT: 5.05V 5V output voltage within spec. 3.3V output voltage within spec. PASS: SET VALUE: 128.62 MEAS VALUE: 129.41 HV output accuracy OK for this sten! PASS: 3.3V MEASUREMENTS after: 55s: 3.39V 5V MEASUREMENT: PASS: 5.03V 5V output voltage within spec. PASS: SET VALUE: MEAS VALUE: 3.3V output voltage within spec 128.62 HV output accuracy OK for this step! 3.3V, 5V and HV output voltage within DASS.

SPEC. STEP 6 DONE WITH NO ERRORS